


Laerdal Silicone Resuscitators

User Guide



www.laerdal.com



Laerdal
helping save lives



Contents

Clinical Indications	4
Important Information	6
Items Included	8
Adult Model	10
Paediatric Model	16
Preterm Model	22
Intake Valve	28
Function Test	29
Clinical Use	36
Cleaning and Disinfection	38
Regulatory Information	42
Specifications	44
Spare Parts and Accessories	48

Clinical Indications

Device Description

The Laerdal Silicone Resuscitator (LSR) is a self-inflating manual resuscitator that is intended for patients requiring total or intermittent ventilatory support.

Indication for Use

The Laerdal Silicone Resuscitator (LSR) is intended for patients requiring total or intermittent ventilatory support. Ventilation is possible with or without supplemental oxygen.

Intended Use

The Laerdal Silicone Resuscitator (LSR) provides positive pressure ventilation and allows spontaneous breathing with a face mask or an artificial airway.

The Laerdal Silicone Resuscitator is available in three sizes:

- The Adult model is intended for patients over 25 kg (44 lb).
- The Paediatric model is intended for patients from 2.5 kg (5.5 lb) to 25 kg (44 lb).
- The Preterm model is intended for patients below 2.5 kg (5.5 lb).

This User Guide applies to all three models of the Laerdal Silicone Resuscitator. For the Masks, refer to the Laerdal Silicone Mask User Guide.

Clinical Indications

Intended Users

The LSR is intended to be used by healthcare professionals trained in delivering ventilatory support and in the use of manual resuscitators.

Clinical Benefits

Positive impact on clinical outcome, by respiratory support that reduces probability of adverse outcomes, such as morbidity and mortality caused by hypoxia.

Clinical Outcome

Desired outcome of ventilation is oxygenation of the patient, often evaluated using SpO_2 , EtCO_2 , blood gas analysis or other method of analysis.

Known Side Effects

Gastric Insufflation
Oxygen Toxicity

Contraindications

No known contraindications for use.



Important Information

Read this User Guide and become familiar with the operation of the product prior to use. Use the product only as described in this User Guide.



Warning and Cautions

A Warning states a condition, hazard, or unsafe practice that can result in serious personal injury or death.

A Caution states a condition, hazard, or unsafe practice that can result in minor personal injury or damage to the manikin.



Notes

Important information about the product or its operation.



Warnings

- *Care should be taken when using the LSR on patients with severe anomalies or when applying other medical devices which may conflict with the mask as mask leakage may occur. If mask face sealing is not possible to achieve consider using alternative airway device.*
- *Care should be taken when using the LSR on patients with severe pulmonary disease or severely immature lungs. Applied pressure should be adjusted and monitored according to the patient's condition. Note that a manometer is not supplied by Laerdal for use with the LSR, but a manometer is possible to connect to the patient port with an appropriate adapter compatible with a ISO 5356-1 connector.*
- *Care should be taken when applying pressure to the mask to avoid facial damage, especially in the case of pediatric patients, infants, pre-terms, patients with severe osteoporosis and geriatric patients.*
- *Care should be taken when using the LSR on patients with*

Important Information



severely congested airways. Consider removing congestion from the oropharyngeal airway. Use of the LSR on patients with severely congested airways may result in a reduction in expected oxygenation.



Cautions

- *The LSR and masks should only be used by persons who have received adequate training in the use of resuscitators.*
- *Resuscitators should not be used with supplemental oxygen where smoking is permitted or when fire, flame, oil or grease is in close proximity.*
- *Resuscitators should not be used in toxic or hazardous atmospheres.*
- *The use of third party products (such as filters and demand valves) with the Laerdal Silicone Resuscitator may affect performance. Please consult with the manufacturer of the third party products to verify compatibility with the LSR and obtain information on possible performance changes.*
- *An oxygen blender is recommended if more precise oxygen concentrations are required, for example for pre-terms.*
- *The use of a PEEP valve (not provided by Laerdal) is recommended in the case that PEEP is indicated for the patient. Note that it is necessary to use the Expiration diverter to attach a PEEP valve.*
- *The LSR and masks are not intended for use in delivery of medications, such as anaesthetic gases.*



Note

Should any serious malfunction, undesirable incident with, or deterioration in the functionality or performance of the device occur, contact Laerdal promptly. The competent authority where the incident took place and/or the device was used should also be notified.

Items Included




Caution

Use of non-Laerdal parts may affect safety and/or performance.

Adult Model
(Cat. No. 87xxxx)



Patient Valve



Adult Ventilation Bag
(1600 ml)



Reusable Oxygen
Reservoir Bag
(2600 ml)



Silicone Mask
(Adult 4-5+)
with Multi-Function
Mask Cover *



Intake
Reservoir Valve

* Some configurations do not include the masks.

Items Included


Paediatric Model (Cat. No. 86xxxx)



Patient Valve with
Pressure Relief Valve



Paediatric Ventilation Bag
(500 ml)



Reusable Oxygen
Reservoir Bag
(600 ml)



Silicone Mask (Child 3-4)
with Multi-Function
Mask Cover *



Intake
Reservoir Valve


Preterm Model (Cat. No. 85xxxx)



Patient Valve with
Pressure Relief Valve



Preterm Ventilation Bag
(240 ml)



Reusable Oxygen
Reservoir Bag
(600 ml)




Silicone Mask *
(Size 00, 0/1, 2)




Intake
Reservoir Valve

Adult Model - Overview




Adult Model - Overview




Adult Model - Overview

Assembly and Disassembly



Adult Model - Overview




Warning


Improper assembly may affect performance. Ensure use of one lip valve. Misassembly with two lip valves may prevent proper patient exhalation.

Adult Model - Overview


Accessories




Adult Model - Overview



Paediatric Model - Overview




Paediatric Model - Overview




Paediatric Model - Overview

Assembly and Disassembly



Paediatric Model - Overview




Warning


Improper assembly may affect performance. Ensure use of one lip valve. Misassembly with two lip valves may prevent proper patient exhalation.

Paediatric Model - Overview


Accessories




Paediatric Model - Overview



Preterm Model - Overview




Preterm Model - Overview




Preterm Model - Overview

Assembly and Disassembly



Preterm Model - Overview




Warning


Improper assembly may affect performance. Ensure use of one lip valve. Misassembly with two lip valves may prevent proper patient exhalation.

Preterm Model - Overview


Accessories



Preterm Model - Overview



Intake Valve Assembly/Disassembly



⚠ Intake valve caps produced prior to 2015 are not compatible with LSRs produced after 2015.



Pre 2015 cap



Post 2015 cap

Function Test

Inspect and test valve function to ensure proper operation of the Laerdal Silicone Resuscitator prior to patient use.

To ensure proper operation, test valve functions after cleaning, disinfection and reassembly.






Caution

If a Laerdal Silicone Resuscitator fails function tests it is to be removed from service and not used. Inspect all parts for damage. Replace any damaged parts if necessary and retest.


Function Test

Testing the Intake Valve



Function Test

Testing the Patient Valve




 **Warning**

Ensure that a single Lip Valve has been installed in the Patient Valve.



Function Test

Testing the Patient Valve Disk Membrane




Function Test

Testing the Oxygen Relief Umbrella Valve



Function Test

Testing the Intake Umbrella Valve



Function Test



Testing the Pressure Relief Valve

Applies to Preterm and Paediatric models.



Caution

Ensure Pressure Relief Valve is functioning prior to use.





Clinical Use

Operating the Laerdal Silicone Resuscitator with face mask:

1. Connect a suitable face mask.
2. Connect to external O₂ source, if applicable.
3. Place mask over face and check for seal.
4. Squeeze the Ventilation Bag in accordance to clinical protocol.
5. Observe patient chest rise during ventilation.
6. Allow patient to exhale.
7. Stop ventilation as required by clinical protocol.



Operating the Laerdal Silicone Resuscitator with advanced airway:



1. Connect to external O₂ source, if applicable.
2. Connect to advanced airway of intubated patients.
3. Squeeze the Ventilation Bag in accordance to clinical protocol.
4. Observe patient chest rise during ventilation.
5. Allow patient to exhale.
6. Stop ventilation as required by clinical protocol.



Warning

Incorrect operation of the resuscitator can be hazardous.



Notes

- *An oxygen tube is not provided with the LSR. The oxygen connector fits oxygen tubes which comply with ISO 13544-2. Fit should be checked prior to use. The oxygen source should be able to be adjusted to provide a flow relevant to the LSR. See tables on pages 46-47 regarding achievable oxygen concentration at varying flows for more information.*
- *Contamination: If the Patient Valve becomes contaminated with vomit during ventilation, disconnect the resuscitator from the patient and clear the Patient Valve as follows:*
 - *Tap the Patient Valve with the patient port against your gloved hand to shake free any contaminant and squeeze the silicone bag to deliver several sharp breaths through the Patient Valve to expel the contaminant.*
 - *If contaminant does not clear; disassemble the Patient Valve and rinse.*

Cleaning and Disinfection

Cautions

- *The resuscitator is not provided sterile. The resuscitator and mask must be cleaned and disinfected prior to initial use.*
- *It is recommended that the highest level of disinfection/sterilization possible is used for patients that may have compromised immune defense, such as a pre-term baby or in the case of outbreaks of highly transmissible pathogens.*
- *If LSR is stored as back-up in an area with potentially high levels of airborne pathogens, it should be considered to store the LSR in an air-tight container to avoid contamination.*

To reduce the risk of cross-contamination, follow these instructions after each use.

Inspection

Carefully inspect all parts for signs of wear or damage. Worn or damaged components must be discarded and replaced with new components.

Disassembly

Disassemble the LSR into individual parts as shown in Assembly and Disassembly section prior to cleaning and disinfecting.

- Separate the Expiration Diverter (if used) into its three parts
- Separate the Patient Valve into its four main parts
- Separate the Intake Reservoir Valve into its six parts
- Do not disassemble the connectors from the Ventilation Bag or the O₂ Reservoir Bag. Do not disassemble the connectors from the Extension Tube, if used.

Cleaning and Disinfection

- Unscrew the Pressure Relief Valve (Preterm and Paediatric models), but do not disassemble any further.

Washing and Rinsing

The LSR and Masks must be cleaned before high-level disinfection or sterilisation.

The LSR and Masks can be manually cleaned, or cleaned in an automatic washer/disinfector.

Manual Cleaning

Rinse parts under cold running water.

Submerge parts in water at 30 - 40 °C (86 - 104 °F).
Ensure that all surfaces are submerged for at least 2 minutes.

Submerge all parts in water at 60 - 70 °C (140 - 158 °F) which contains dish washing detergent.

Thoroughly clean all surfaces using a brush as necessary.

Rinse all components in detergent-free water at 30 - 40 °C (86 - 104 °F).

Dry the components thoroughly. Inspect all components to confirm that they are clean and dry. If parts are worn or damaged, discard them.

Automatic Cleaning *(applies to all parts except O₂ Reservoir Bags)*

Washer/Disinfector

Place parts in wire baskets.

Cycle 1: 90 - 95 °C (194 - 203 °F) for more than 12 seconds.

Total process time: approx. 52 min.

Cycle 2: Use a Non-enzymatic alkaline detergent containing 2 - 5% NaOH.



Caution

Do not use rinsing and drying agents.

Cleaning and Disinfection

To obtain high-level disinfection/sterilisation of the LSR and Masks, follow one of these methods.

Sterilisation/High-level Disinfection			
Method	Process Parameters		Post-Treatment
	Temperature / Concentration	Exposure time	
Sterilisation (applies to all parts except the O ₂ Reservoir Bags which do not withstand high temperatures)			
Steam Autoclaving (prevacuum-pulse)	Autoclave at 134 - 137 °C (273 - 279 °F)	3 min (+30s)	Allow parts to cool and dry
High-level Disinfection (applies to all parts)			
Cidex OPA (orthophthalaldehyde)	0.55% solution	60 min	Remove traces of disinfectant by rinsing in warm tap water; 30 - 40 °C (86 - 104 °F), for at least 2 mins. Dry the components thoroughly.
Sodium Hypochlorite	0.5% solution	20 min	

Reassembly

Reassemble LSR as shown in Assembly/Disassembly section.



Note

Perform function test after assembly and before patient use.

Cleaning and Disinfection

Warning

Disposable Oxygen Reservoir Bag (870702)

Designed for single patient use only. Do not reuse. Reuse will lead to risk of cross contamination. Laerdal is not responsible for any consequences of reuse.

Cautions

- The resuscitator components must be cleaned and disinfected before next patient use.*
- The use of cleaning and disinfection procedures not described in this section may have adverse effects on the LSR material and/or performance and may not be effective for disinfecting the LSR.*
- The hard plastic components of the resuscitator and the mask cover are incompatible with polar solvents such as ethanol and isopropyl alcohol.*
- Improper assembly of the LSR after reprocessing may affect performance.*
- Accessories used for storing the LSR are not compatible with sodium hypochlorite.*

Regulatory Information

Laerdal Silicone Resuscitator meets the following Standards:

- EN 1789:2020
- ISO 10651-4:2002

When used in accordance with ISO 10651-4 the following resuscitator size recommendation applies: Adult for patients over 20 kg (44 lb), Paediatric for patients from 2.5 kg (5.5 lb) to 20 kg (44 lb) and Preterm for patients below 2.5 kg (5.5 lb).

When used to deliver tidal volumes as recommended by the AHA Guidelines 2010, the following applies: Adult for patients over 25 kg (55 lb), Paediatric for patients from 2.5 kg (5.5 lb) to 25 kg (55 lb) and Preterm for patients below 2.5 kg (5.5 lb).

Regulatory Information

Symbol Glossary	
	Medical Device
	This medical device complies with the general safety and performance requirements of Regulation (EU) 2017/745 for medical devices.
	Caution: Federal law restricts this device to sale by, or on the order of a physician (US).
	Not made with natural rubber latex
	Single-use symbol

Specifications

Conditions	
Operating Conditions	Temperature: -18 °C to 60 °C (0 °F to 140 °F) Humidity: 15% to 95% rH
Storage Conditions	Temperature: -40 °C to 70 °C (-40 °F to 158 °F) Humidity: 15% to 95% rH
Lifetime Parameters	
Shelf-life	5 years
Expected Service Life	100 cycles of reprocessing
Resistance	
Expiratory resistance	Approximately 2.6 cm H ₂ O Measured with airflow of 50 lpm
Inspiratory resistance	With O ₂ Reservoir: approx. 4.2 cm H ₂ O Without O ₂ Reservoir: approx. 3.1 cm H ₂ O Measured with airflow of 50 lpm
Attainable delivery volume	
Adult	Approximately 800 ml
Paediatric	Approximately 320 ml
Preterm	Approximately 150 ml
Test conditions	Compliance 0.02 l/cm H ₂ O, Resistance 20 cm H ₂ O/l/s
No leakage	Pressure Relief Valve overridden
Dead space of Patient Valve	Approximately 7 ml for all models

Specifications

Material Chart

Resuscitator		Accessories	
Parts	Materials	Parts	Materials
Mask	PSU, Silicone	Expiration Diverter	PSU, Silicone
Patient Valve (w/ Pressure Relief Valve)	PSU, Silicone (PPSU, Steel)	Silicone Extension Tube	PSU, Silicone, Viton
Ventilation Bag	PSU, Silicone, Viton	Hanging loop	Silicone
Intake Valve	PSU, Silicone	Wall Bracket	POM
Oxygen Reservoir	PC, PTFE, PVC	Wall Mount	ABS
		Display Case	ABS, PA, PP, Steel
		Disposable Oxygen Reservoir	PVC, PC

Specifications

Adult Model

Ventilation Bag volume: 1600 ml.

Reservoir Bag volume: 2600 ml

Weight: Approximately 370 g

Dimensions: Approximately 370 mm x 132 mm x 132 mm

Dimension Display Case: W 291/326 mm x L 362 mm x H 136 mm

Dimension Compact Case: W 163/189 mm x L 237 mm x H 150 mm

Delivered O₂ concentrations under various test conditions

O ₂ flow (lpm)	Tidal volume (ml) x bag cycling rate per minute.					
	O ₂ -concentrations (%) using O ₂ Reservoir			(without O ₂ Reservoir)		
	400 x 12	400 x 24	600 x 12	600 x 24	1000 x 12	1000 x 24
3	74 (38)	51 (39)	58 (34)	40 (34)	44 (33)	33 (30)
8	100 (44)	100 (44)	100 (40)	68 (40)	78 (38)	51 (34)
15	100 (51)	100 (50)	100 (47)	100 (47)	100 (42)	75 (36)

Paediatric Model

Ventilation Bag volume: 500 ml.

Reservoir Bag volume: 600 ml

Weight: Approximately 230 g

Dimensions: Approximately 300 mm x 88 mm x 93 mm

Dimension Display Case: W 291/326 mm x L 362 mm x H 110 mm

Dimension Compact Case: W 163/189 mm x L 237 mm x H 150 mm

Delivered O₂ concentrations under various test conditions

O ₂ flow (lpm)	Tidal volume (ml) x bag cycling rate per minute.					
	O ₂ -concentrations (%) using O ₂ Reservoir			(without O ₂ Reservoir)		
	20 x 40	20 x 60	150 x 20	150 x 30	300 x 12	300 x 24
3	100 (97)	100 (97)	98 (56)	78 (57)	85 (48)	56 (46)

Specifications

8	100 (100)	100 (100)	100 (70)	100 (70)	100 (58)	100 (57)
15	100 (100)	100 (100)	100 (82)	100 (83)	100 (71)	100 (70)

Preterm Model

Ventilation Bag volume: 240 ml.

Reservoir Bag volume: 600 ml

Weight: Approximately 200 g

Dimensions: Approximately 280 mm x 72 mm x 85 mm

Dimension Display Case: W 291/326 mm x L 362 mm x H 110 mm

Dimension Compact Case: W 163/189 mm x L 237 mm x H 150 mm

Delivered O₂ concentrations under various test conditions

O ₂ flow (lpm)	Tidal volume (ml) x bag cycling rate per minute. O ₂ -concentrations (%) using O ₂ Reservoir (without O ₂ Reservoir)					
	20 x 40	20 x 60				
3	100 (98)	100 (97)				
8	100 (100)	100 (100)				
15	100 (100)	100 (100)				

Spare Parts and Accessories

Accessories

Catalogue #	Description
511700	LSR Wall Bracket
521100	Wall Mount for Paediatric and Preterm Display Case
572000	Wall Mount for Adult Display Case
850500	Expiration Diverter (OD 30 mm)
860300	Display Case, Paediatric
870120	LSR Hanging Loop
870600	LSR Display Case Complete Adult
870702	Disposable O ₂ Reservoir
871000	Silicone Extension Tube

Spare Parts

Catalogue #	Description
510103	Cap for LSR Intake Valve, pack of 3
510404	LSR Intake Membranes, pack of 10
531901	LSR O ₂ Reservoir 2.6 l
531906	LSR O ₂ Reservoir 2.6 l
540103	LSR Lip Valve
540105	LSR Disk Membranes, pack of 10
551901	LSR O ₂ Reservoir, 0.6 l

Spare Parts and Accessories

551906	Oxygen Reservoir Bag 0.6 l, pack of 50
560200	LSR Patient Valve
850150	Preterm Bag, 240ml
851103	LSR Lock Clips, pack of 10
851250	Patient Valve with 35cm H ₂ O Pressure Relief Valve
851252	Pressure Relief Valve 35 cm H ₂ O
851350	Patient Valve with 35cm H ₂ O with Pressure Relief Valve and Lock Clip
860150	Paediatric Bag 500ml
870150	Adult Bag 1600ml
871950	Umbrella Valves, pack of 2
875400xx	Intake Reservoir Valve

Masks - Main Products

Catalogue #	Description
851500xx	LSR Silicone Mask no. 00
851600xx	LSR Silicone Mask no. 0/1
851700xx	LSR Silicone Mask no. 2
860220xx	Child Silicone Mask 3-4 with Multi Function Mask Cover
860221	Child Silicone Mask 3-4 without Multi Function Mask Cover
870220xx	Adult Silicone Mask 4-5+ with Multi Function Mask Cover

Spare Parts and Accessories

870221	Adult Silicone Mask 4-5+ without Multi Function Mask Cover
872220	Adult & Child Silicone Mask with Multi Function Mask Covers



Note

Catalog numbers ending with xx denotes local language configurations

Masks - Spare Parts/Accessories

Catalogue #	Description
865200	Multi Function Mask Cover for Mask 3-4
875200	Multi Function Mask Cover for Mask 4-5+

For latest version of Spare Parts and Accessories, visit www.laerdal.com.

Warranty

Refer to the Laerdal Global Warranty for terms and conditions.
For more information visit www.laerdal.com.



Laerdal® is a registered trademark
of Laerdal Medical AS.

© 2022 Laerdal Medical AS. All rights reserved.



Manufactured in China for Laerdal Medical AS
P.O. Box 377, Tanke Svilandsgate 30,
4002 Stavanger, Norway
T: (+47) 51 51 17 00

Printed in China

20-19627 Rev A

www.laerdal.com



Laerdal
helping save lives